

DNL Test Results

File Root Name - 2010051915350xDB8EC9
Nominal Value = 241.6422
Minimum Bin Value at Bin 201827 83 -0.6565
Maximum Bin Value at Bin 228392 354 0.4650
Channel - 1 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 239.4022
Minimum Bin Value at Bin 241439 94 -0.6074
Maximum Bin Value at Bin 222248 328 0.3701
Channel - 2 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 236.8372
Minimum Bin Value at Bin 203859 80 -0.6622
Maximum Bin Value at Bin 191764 335 0.4145
Channel - 3 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 242.5546
Minimum Bin Value at Bin 4639 92 -0.6207
Maximum Bin Value at Bin 205396 318 0.3110
Channel - 4 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 255.9782
Minimum Bin Value at Bin 229671 81 -0.6836
Maximum Bin Value at Bin 260648 366 0.4298
Channel - 5 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 259.8544
Minimum Bin Value at Bin 213187 93 -0.6421
Maximum Bin Value at Bin 165412 350 0.3469
Channel - 6 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 241.0455
Minimum Bin Value at Bin 197803 99 -0.5893
Maximum Bin Value at Bin 224296 346 0.4354
Channel - 7 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 239.9499
Minimum Bin Value at Bin 248851 89 -0.6291
Maximum Bin Value at Bin 258600 354 0.4753
Channel - 8 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 242.6058
Minimum Bin Value at Bin 252583 84 -0.6538
Maximum Bin Value at Bin 246824 343 0.4138
Channel - 9 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 240.3545
Minimum Bin Value at Bin 215743 90 -0.6256
Maximum Bin Value at Bin 215336 309 0.2856
Channel - 10 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 242.7838
Minimum Bin Value at Bin 241043 92 -0.6211
Maximum Bin Value at Bin 168996 335 0.3798
Channel - 11 DNL Test PASSED

File Root Name - 2010051915350xDB8EC9
Nominal Value = 243.2876
Minimum Bin Value at Bin 247207 82 -0.6630
Maximum Bin Value at Bin 187684 323 0.3276
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____