

DNL Test Results

File Root Name - 2010081614220xDB8EE6
Nominal Value = 275.1219
Minimum Bin Value at Bin 218191 136 -0.5057
Maximum Bin Value at Bin 173094 476 0.7301
Channel - 1 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 277.5985
Minimum Bin Value at Bin 155587 136 -0.5101
Maximum Bin Value at Bin 104740 475 0.7111
Channel - 2 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 256.7658
Minimum Bin Value at Bin 196739 121 -0.5288
Maximum Bin Value at Bin 146468 419 0.6318
Channel - 3 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 278.6004
Minimum Bin Value at Bin 24179 145 -0.4795
Maximum Bin Value at Bin 212006 495 0.7767
Channel - 4 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 276.0986
Minimum Bin Value at Bin 251303 126 -0.5436
Maximum Bin Value at Bin 105508 455 0.6480
Channel - 5 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 270.4568
Minimum Bin Value at Bin 145603 130 -0.5193
Maximum Bin Value at Bin 183334 424 0.5677
Channel - 6 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 264.6969
Minimum Bin Value at Bin 261159 114 -0.5693
Maximum Bin Value at Bin 247336 418 0.5792
Channel - 7 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 277.7880
Minimum Bin Value at Bin 261543 137 -0.5068
Maximum Bin Value at Bin 158756 456 0.6415
Channel - 8 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 265.2813
Minimum Bin Value at Bin 240915 105 -0.6042
Maximum Bin Value at Bin 230438 455 0.7152
Channel - 9 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 276.9722
Minimum Bin Value at Bin 234515 144 -0.4801
Maximum Bin Value at Bin 224296 430 0.5525
Channel - 10 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 270.1526
Minimum Bin Value at Bin 189075 128 -0.5262
Maximum Bin Value at Bin 235048 422 0.5621
Channel - 11 DNL Test PASSED

File Root Name - 2010081614220xDB8EE6
Nominal Value = 278.6275
Minimum Bin Value at Bin 201803 139 -0.5011
Maximum Bin Value at Bin 214056 485 0.7407
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____