

DNL Test Results

File Root Name - 2010052515200xDB8F9B
Nominal Value = 240.9612
Minimum Bin Value at Bin 235523 87 -0.6389
Maximum Bin Value at Bin 260648 332 0.3778
Channel - 1 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 240.8636
Minimum Bin Value at Bin 156903 100 -0.5848
Maximum Bin Value at Bin 86820 368 0.5278
Channel - 2 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 240.1443
Minimum Bin Value at Bin 215699 91 -0.6211
Maximum Bin Value at Bin 134180 361 0.5033
Channel - 3 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 241.6286
Minimum Bin Value at Bin 252339 75 -0.6896
Maximum Bin Value at Bin 180900 335 0.3864
Channel - 4 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 261.5603
Minimum Bin Value at Bin 209987 91 -0.6521
Maximum Bin Value at Bin 245032 341 0.3037
Channel - 5 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 259.6939
Minimum Bin Value at Bin 261031 87 -0.6650
Maximum Bin Value at Bin 171812 336 0.2938
Channel - 6 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 236.9823
Minimum Bin Value at Bin 204099 78 -0.6709
Maximum Bin Value at Bin 218152 322 0.3588
Channel - 7 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 240.2356
Minimum Bin Value at Bin 254375 90 -0.6254
Maximum Bin Value at Bin 174628 331 0.3778
Channel - 8 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 239.8401
Minimum Bin Value at Bin 215207 86 -0.6414
Maximum Bin Value at Bin 199460 334 0.3926
Channel - 9 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 232.8152
Minimum Bin Value at Bin 256551 64 -0.7251
Maximum Bin Value at Bin 199972 311 0.3358
Channel - 10 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 241.1076
Minimum Bin Value at Bin 237067 86 -0.6433
Maximum Bin Value at Bin 255018 320 0.3272
Channel - 11 DNL Test PASSED

File Root Name - 2010052515200xDB8F9B
Nominal Value = 242.7231
Minimum Bin Value at Bin 211523 84 -0.6539
Maximum Bin Value at Bin 261162 398 0.6397
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____