

DNL Test Results

File Root Name - 2010070810380xDB6F97
Nominal Value = 250.5055
Minimum Bin Value at Bin 211167 102 -0.5928
Maximum Bin Value at Bin 228392 410 0.6367
Channel - 1 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 251.3151
Minimum Bin Value at Bin 257891 107 -0.5742
Maximum Bin Value at Bin 194084 384 0.5280
Channel - 2 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 248.4331
Minimum Bin Value at Bin 260631 95 -0.6176
Maximum Bin Value at Bin 230440 371 0.4934
Channel - 3 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 249.6316
Minimum Bin Value at Bin 251407 102 -0.5914
Maximum Bin Value at Bin 220200 385 0.5423
Channel - 4 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 273.8740
Minimum Bin Value at Bin 240579 101 -0.6312
Maximum Bin Value at Bin 180516 412 0.5043
Channel - 5 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 271.0923
Minimum Bin Value at Bin 211719 105 -0.6127
Maximum Bin Value at Bin 255018 411 0.5161
Channel - 6 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 252.0793
Minimum Bin Value at Bin 196103 101 -0.5993
Maximum Bin Value at Bin 220200 405 0.6066
Channel - 7 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 246.7980
Minimum Bin Value at Bin 217747 96 -0.6110
Maximum Bin Value at Bin 226346 401 0.6248
Channel - 8 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 248.4806
Minimum Bin Value at Bin 255507 89 -0.6418
Maximum Bin Value at Bin 261672 375 0.5092
Channel - 9 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 250.6814
Minimum Bin Value at Bin 226255 99 -0.6051
Maximum Bin Value at Bin 78884 379 0.5119
Channel - 10 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 250.4902
Minimum Bin Value at Bin 241703 89 -0.6447
Maximum Bin Value at Bin 97316 384 0.5330
Channel - 11 DNL Test PASSED

File Root Name - 2010070810380xDB6F97
Nominal Value = 253.5299
Minimum Bin Value at Bin 213119 107 -0.5780
Maximum Bin Value at Bin 101412 404 0.5935
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____