

#### DNL Test Results

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 269.0287  
Minimum Bin Value at Bin 187347 129 -0.5205  
Maximum Bin Value at Bin 185382 474 0.7619  
Channel - 1 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 272.5334  
Minimum Bin Value at Bin 91211 130 -0.5230  
Maximum Bin Value at Bin 166692 439 0.6108  
Channel - 2 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 267.6601  
Minimum Bin Value at Bin 220595 119 -0.5554  
Maximum Bin Value at Bin 155428 464 0.7335  
Channel - 3 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 276.0607  
Minimum Bin Value at Bin 175179 142 -0.4856  
Maximum Bin Value at Bin 230440 456 0.6518  
Channel - 4 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 256.4302  
Minimum Bin Value at Bin 187467 112 -0.5632  
Maximum Bin Value at Bin 174372 422 0.6457  
Channel - 5 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 273.9119  
Minimum Bin Value at Bin 247491 123 -0.5510  
Maximum Bin Value at Bin 114980 468 0.7086  
Channel - 6 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 277.5337  
Minimum Bin Value at Bin 238659 134 -0.5172  
Maximum Bin Value at Bin 250922 488 0.7583  
Channel - 7 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 277.7829  
Minimum Bin Value at Bin 257103 126 -0.5464  
Maximum Bin Value at Bin 163876 455 0.6380  
Channel - 8 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 277.7277  
Minimum Bin Value at Bin 250931 138 -0.5031  
Maximum Bin Value at Bin 250922 488 0.7571  
Channel - 9 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 261.4655  
Minimum Bin Value at Bin 185075 118 -0.5487  
Maximum Bin Value at Bin 119844 408 0.5604  
Channel - 10 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 278.6880  
Minimum Bin Value at Bin 217731 127 -0.5443  
Maximum Bin Value at Bin 109604 490 0.7582  
Channel - 11 DNL Test PASSED

File Root Name - 2010081615550xDB8EEC  
Nominal Value = 264.8376  
Minimum Bin Value at Bin 255647 127 -0.5205  
Maximum Bin Value at Bin 245288 427 0.6123  
Channel - 12 DNL Test PASSED

\_\_\_\_\_All 12 Channels Pass DNL Test\_\_\_\_\_