

#### DNL Test Results

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 276.2679  
Minimum Bin Value at Bin 251027 144 -0.4788  
Maximum Bin Value at Bin 175142 458 0.6578  
Channel - 1 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 275.9821  
Minimum Bin Value at Bin 196403 140 -0.4927  
Maximum Bin Value at Bin 213028 459 0.6632  
Channel - 2 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 276.5422  
Minimum Bin Value at Bin 255399 134 -0.5154  
Maximum Bin Value at Bin 228392 517 0.8695  
Channel - 3 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 278.5953  
Minimum Bin Value at Bin 245555 144 -0.4831  
Maximum Bin Value at Bin 198436 460 0.6511  
Channel - 4 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 277.7933  
Minimum Bin Value at Bin 214911 144 -0.4816  
Maximum Bin Value at Bin 218152 475 0.7099  
Channel - 5 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 276.2750  
Minimum Bin Value at Bin 193475 135 -0.5114  
Maximum Bin Value at Bin 119844 494 0.7881  
Channel - 6 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 278.1860  
Minimum Bin Value at Bin 244583 129 -0.5363  
Maximum Bin Value at Bin 259624 443 0.5925  
Channel - 7 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 277.4196  
Minimum Bin Value at Bin 221043 133 -0.5206  
Maximum Bin Value at Bin 151332 461 0.6617  
Channel - 8 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 275.3602  
Minimum Bin Value at Bin 225763 128 -0.5352  
Maximum Bin Value at Bin 257320 436 0.5834  
Channel - 9 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 280.0177  
Minimum Bin Value at Bin 226371 129 -0.5393  
Maximum Bin Value at Bin 109604 492 0.7570  
Channel - 10 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 276.9530  
Minimum Bin Value at Bin 256679 116 -0.5812  
Maximum Bin Value at Bin 259114 486 0.7548  
Channel - 11 DNL Test PASSED

File Root Name - 2010081615150xDB6FA8  
Nominal Value = 279.0494  
Minimum Bin Value at Bin 220611 141 -0.4947  
Maximum Bin Value at Bin 244262 449 0.6090  
Channel - 12 DNL Test PASSED

\_\_\_\_\_All 12 Channels Pass DNL Test\_\_\_\_\_