

DNL Test Results

File Root Name - 2010081707380xDB8F7D
Nominal Value = 277.0000
Minimum Bin Value at Bin 223091 143 -0.4838
Maximum Bin Value at Bin 259114 464 0.6751
Channel - 1 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 277.5793
Minimum Bin Value at Bin 149843 145 -0.4776
Maximum Bin Value at Bin 175396 486 0.7509
Channel - 2 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 276.7597
Minimum Bin Value at Bin 212371 132 -0.5231
Maximum Bin Value at Bin 146468 482 0.7416
Channel - 3 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 277.3324
Minimum Bin Value at Bin 223399 116 -0.5817
Maximum Bin Value at Bin 113700 490 0.7668
Channel - 4 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 275.6860
Minimum Bin Value at Bin 156403 138 -0.4994
Maximum Bin Value at Bin 224296 462 0.6758
Channel - 5 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 279.5209
Minimum Bin Value at Bin 202291 138 -0.5063
Maximum Bin Value at Bin 240680 498 0.7816
Channel - 6 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 280.3917
Minimum Bin Value at Bin 208883 138 -0.5078
Maximum Bin Value at Bin 255018 475 0.6941
Channel - 7 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 276.2496
Minimum Bin Value at Bin 252471 146 -0.4715
Maximum Bin Value at Bin 261162 462 0.6724
Channel - 8 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 278.0885
Minimum Bin Value at Bin 227799 139 -0.5002
Maximum Bin Value at Bin 218152 505 0.8160
Channel - 9 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 275.9762
Minimum Bin Value at Bin 227907 131 -0.5253
Maximum Bin Value at Bin 119844 463 0.6777
Channel - 10 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 276.6457
Minimum Bin Value at Bin 250595 144 -0.4795
Maximum Bin Value at Bin 245288 455 0.6447
Channel - 11 DNL Test PASSED

File Root Name - 2010081707380xDB8F7D
Nominal Value = 279.0209
Minimum Bin Value at Bin 217795 140 -0.4982
Maximum Bin Value at Bin 261162 521 0.8672
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____