

DNL Test Results

File Root Name - 2010052010560xDB6FA0
Nominal Value = 242.6563
Minimum Bin Value at Bin 229491 80 -0.6703
Maximum Bin Value at Bin 3875 363 0.4959
Channel - 1 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 239.4882
Minimum Bin Value at Bin 246355 82 -0.6576
Maximum Bin Value at Bin 226344 355 0.4823
Channel - 2 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 241.7224
Minimum Bin Value at Bin 232455 79 -0.6732
Maximum Bin Value at Bin 217700 341 0.4107
Channel - 3 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 242.4232
Minimum Bin Value at Bin 167427 97 -0.5999
Maximum Bin Value at Bin 20002 331 0.3654
Channel - 4 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 260.5739
Minimum Bin Value at Bin 232103 82 -0.6853
Maximum Bin Value at Bin 258600 341 0.3086
Channel - 5 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 260.5596
Minimum Bin Value at Bin 206915 83 -0.6815
Maximum Bin Value at Bin 181540 356 0.3663
Channel - 6 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 240.1024
Minimum Bin Value at Bin 246775 74 -0.6918
Maximum Bin Value at Bin 218152 355 0.4785
Channel - 7 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 239.1867
Minimum Bin Value at Bin 224867 90 -0.6237
Maximum Bin Value at Bin 226344 341 0.4257
Channel - 8 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 242.0777
Minimum Bin Value at Bin 236159 74 -0.6943
Maximum Bin Value at Bin 242728 358 0.4789
Channel - 9 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 241.7675
Minimum Bin Value at Bin 257191 88 -0.6360
Maximum Bin Value at Bin 234536 329 0.3608
Channel - 10 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 241.9130
Minimum Bin Value at Bin 241923 80 -0.6693
Maximum Bin Value at Bin 219588 341 0.4096
Channel - 11 DNL Test PASSED

File Root Name - 2010052010560xDB6FA0
Nominal Value = 242.6682
Minimum Bin Value at Bin 244675 93 -0.6168
Maximum Bin Value at Bin 191528 373 0.5371
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____