

DNL Test Results

File Root Name - 2010070810060xDB6FB0
Nominal Value = 249.7357
Minimum Bin Value at Bin 228243 96 -0.6156
Maximum Bin Value at Bin 240678 403 0.6137
Channel - 1 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 249.3517
Minimum Bin Value at Bin 213955 108 -0.5669
Maximum Bin Value at Bin 250408 374 0.4999
Channel - 2 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 251.6088
Minimum Bin Value at Bin 238039 101 -0.5986
Maximum Bin Value at Bin 246824 418 0.6613
Channel - 3 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 250.7361
Minimum Bin Value at Bin 225331 97 -0.6131
Maximum Bin Value at Bin 111652 380 0.5155
Channel - 4 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 272.1070
Minimum Bin Value at Bin 142659 108 -0.6031
Maximum Bin Value at Bin 62500 462 0.6979
Channel - 5 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 272.3940
Minimum Bin Value at Bin 258975 101 -0.6292
Maximum Bin Value at Bin 136996 423 0.5529
Channel - 6 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 249.2122
Minimum Bin Value at Bin 221351 92 -0.6308
Maximum Bin Value at Bin 166948 406 0.6291
Channel - 7 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 250.9157
Minimum Bin Value at Bin 208707 94 -0.6254
Maximum Bin Value at Bin 215844 347 0.3829
Channel - 8 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 250.7216
Minimum Bin Value at Bin 168915 106 -0.5772
Maximum Bin Value at Bin 66596 361 0.4398
Channel - 9 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 241.5914
Minimum Bin Value at Bin 211351 91 -0.6233
Maximum Bin Value at Bin 146468 374 0.5481
Channel - 10 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 251.0457
Minimum Bin Value at Bin 203639 104 -0.5857
Maximum Bin Value at Bin 230440 379 0.5097
Channel - 11 DNL Test PASSED

File Root Name - 2010070810060xDB6FB0
Nominal Value = 253.3408
Minimum Bin Value at Bin 149907 95 -0.6250
Maximum Bin Value at Bin 186404 356 0.4052
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____