

DNL Test Results

File Root Name - 2010061514380xDB6FB8
Nominal Value = 248.7139
Minimum Bin Value at Bin 158447 109 -0.5617
Maximum Bin Value at Bin 259114 392 0.5761
Channel - 1 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 248.9700
Minimum Bin Value at Bin 256727 97 -0.6104
Maximum Bin Value at Bin 123940 398 0.5986
Channel - 2 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 248.4208
Minimum Bin Value at Bin 128403 109 -0.5612
Maximum Bin Value at Bin 156964 340 0.3686
Channel - 3 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 251.7974
Minimum Bin Value at Bin 230343 100 -0.6029
Maximum Bin Value at Bin 226344 396 0.5727
Channel - 4 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 272.2731
Minimum Bin Value at Bin 218983 94 -0.6548
Maximum Bin Value at Bin 248872 382 0.4030
Channel - 5 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 273.4343
Minimum Bin Value at Bin 210727 100 -0.6343
Maximum Bin Value at Bin 230440 378 0.3824
Channel - 6 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 248.1681
Minimum Bin Value at Bin 251895 88 -0.6454
Maximum Bin Value at Bin 259368 372 0.4990
Channel - 7 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 249.2832
Minimum Bin Value at Bin 253943 94 -0.6229
Maximum Bin Value at Bin 208932 385 0.5444
Channel - 8 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 249.1097
Minimum Bin Value at Bin 250355 97 -0.6106
Maximum Bin Value at Bin 261160 411 0.6499
Channel - 9 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 249.8481
Minimum Bin Value at Bin 230115 103 -0.5877
Maximum Bin Value at Bin 226344 398 0.5930
Channel - 10 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 250.0319
Minimum Bin Value at Bin 211411 105 -0.5801
Maximum Bin Value at Bin 250922 392 0.5678
Channel - 11 DNL Test PASSED

File Root Name - 2010061514380xDB6FB8
Nominal Value = 250.9517
Minimum Bin Value at Bin 229647 97 -0.6135
Maximum Bin Value at Bin 218152 422 0.6816
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____