

DNL Test Results

File Root Name - 2010081711050xDB8F85
Nominal Value = 274.6786
Minimum Bin Value at Bin 242543 139 -0.4940
Maximum Bin Value at Bin 222248 432 0.5727
Channel - 1 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 276.8546
Minimum Bin Value at Bin 233859 136 -0.5088
Maximum Bin Value at Bin 255018 507 0.8313
Channel - 2 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 275.5231
Minimum Bin Value at Bin 228003 128 -0.5354
Maximum Bin Value at Bin 168228 496 0.8002
Channel - 3 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 278.4099
Minimum Bin Value at Bin 205891 131 -0.5295
Maximum Bin Value at Bin 142372 479 0.7205
Channel - 4 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 277.6287
Minimum Bin Value at Bin 252583 133 -0.5209
Maximum Bin Value at Bin 180004 478 0.7217
Channel - 5 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 279.9806
Minimum Bin Value at Bin 205619 136 -0.5143
Maximum Bin Value at Bin 146468 500 0.7858
Channel - 6 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 280.0364
Minimum Bin Value at Bin 232455 128 -0.5429
Maximum Bin Value at Bin 261716 475 0.6962
Channel - 7 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 279.6252
Minimum Bin Value at Bin 224083 132 -0.5279
Maximum Bin Value at Bin 226344 481 0.7202
Channel - 8 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 275.2615
Minimum Bin Value at Bin 228931 122 -0.5568
Maximum Bin Value at Bin 172836 461 0.6748
Channel - 9 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 279.2927
Minimum Bin Value at Bin 243447 129 -0.5381
Maximum Bin Value at Bin 242728 494 0.7688
Channel - 10 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 274.9003
Minimum Bin Value at Bin 248359 126 -0.5417
Maximum Bin Value at Bin 199332 492 0.7897
Channel - 11 DNL Test PASSED

File Root Name - 2010081711050xDB8F85
Nominal Value = 278.6764
Minimum Bin Value at Bin 167651 130 -0.5335
Maximum Bin Value at Bin 246824 515 0.8480
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____