

DNL Test Results

File Root Name - 2010052514010xDB8F75
Nominal Value = 239.3273
Minimum Bin Value at Bin 212035 87 -0.6365
Maximum Bin Value at Bin 216104 361 0.5084
Channel - 1 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 239.0202
Minimum Bin Value at Bin 211139 82 -0.6569
Maximum Bin Value at Bin 207268 340 0.4225
Channel - 2 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 238.3185
Minimum Bin Value at Bin 226855 84 -0.6475
Maximum Bin Value at Bin 192804 328 0.3763
Channel - 3 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 240.4228
Minimum Bin Value at Bin 259059 89 -0.6298
Maximum Bin Value at Bin 157988 335 0.3934
Channel - 4 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 260.7150
Minimum Bin Value at Bin 256855 74 -0.7162
Maximum Bin Value at Bin 218152 351 0.3463
Channel - 5 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 260.0233
Minimum Bin Value at Bin 245107 86 -0.6693
Maximum Bin Value at Bin 210724 366 0.4076
Channel - 6 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 240.1451
Minimum Bin Value at Bin 222407 79 -0.6710
Maximum Bin Value at Bin 228392 352 0.4658
Channel - 7 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 239.3927
Minimum Bin Value at Bin 245059 90 -0.6240
Maximum Bin Value at Bin 105508 334 0.3952
Channel - 8 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 238.7224
Minimum Bin Value at Bin 212691 83 -0.6523
Maximum Bin Value at Bin 226344 330 0.3824
Channel - 9 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 239.0646
Minimum Bin Value at Bin 194243 90 -0.6235
Maximum Bin Value at Bin 228648 314 0.3135
Channel - 10 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 239.9732
Minimum Bin Value at Bin 220579 81 -0.6625
Maximum Bin Value at Bin 246824 336 0.4002
Channel - 11 DNL Test PASSED

File Root Name - 2010052514010xDB8F75
Nominal Value = 241.9257
Minimum Bin Value at Bin 226371 83 -0.6569
Maximum Bin Value at Bin 208164 368 0.5211
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____