

DNL Test Results

File Root Name - 2010052507390xDB8EC0
Nominal Value = 238.3088
Minimum Bin Value at Bin 237719 81 -0.6601
Maximum Bin Value at Bin 244776 319 0.3386
Channel - 1 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 236.9359
Minimum Bin Value at Bin 209623 87 -0.6328
Maximum Bin Value at Bin 220200 335 0.4139
Channel - 2 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 239.6427
Minimum Bin Value at Bin 218035 85 -0.6453
Maximum Bin Value at Bin 220200 315 0.3145
Channel - 3 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 241.2622
Minimum Bin Value at Bin 251875 79 -0.6726
Maximum Bin Value at Bin 224296 341 0.4134
Channel - 4 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 259.3984
Minimum Bin Value at Bin 205075 83 -0.6800
Maximum Bin Value at Bin 224296 338 0.3030
Channel - 5 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 259.8038
Minimum Bin Value at Bin 206659 88 -0.6613
Maximum Bin Value at Bin 154916 342 0.3164
Channel - 6 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 241.4633
Minimum Bin Value at Bin 254631 71 -0.7060
Maximum Bin Value at Bin 152612 374 0.5489
Channel - 7 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 241.3466
Minimum Bin Value at Bin 254003 94 -0.6105
Maximum Bin Value at Bin 207910 342 0.4170
Channel - 8 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 240.0246
Minimum Bin Value at Bin 185299 94 -0.6084
Maximum Bin Value at Bin 257064 349 0.4540
Channel - 9 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 241.3032
Minimum Bin Value at Bin 226883 79 -0.6726
Maximum Bin Value at Bin 261162 372 0.5416
Channel - 10 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 242.0710
Minimum Bin Value at Bin 236999 85 -0.6489
Maximum Bin Value at Bin 255018 342 0.4128
Channel - 11 DNL Test PASSED

File Root Name - 2010052507390xDB8EC0
Nominal Value = 241.1944
Minimum Bin Value at Bin 1568 88 -0.6351
Maximum Bin Value at Bin 173348 328 0.3599
Channel - 12 DNL Test PASSED

_____All 12 Channels Pass DNL Test_____